



169 Lexington Court, Unit i
Waterloo, ON Canada, N2J 4R9
Toll Free: (877) 746-4764
Local: (519) 746-4761
Email: info@indigoinstruments.com

CERTIFICATION OF ANALYSIS

Product Code: [33810-B8](#)

Product Description: Red Litmus 8x10" Sheet

Application: Measuring pH of water samples in the range 3.0-6.0 in 0.5 pH unit increments

The papers used to produce these are impregnated with solutions of acid/base indicators. There are no established health hazards, and the test papers are stable if stored properly (protect from sunlight and excessive heat until use).

The papers are tested using certified buffer solutions.

For precision, master samples are pulled and checked for positive and consistent reaction to Acidic buffers.