



169 Lexington Court, Unit i
Waterloo, ON Canada N2J 4R9
Toll Free: (877) 746-4764
Local: (519) 746-4761
Email: info@indigoinstruments.com

Stability Statement

Part Number: [33810-Phe](#)

Product Description: Ammonia Leak Detection Strips (Phenolphthalein).

The test papers should be stored in closed vials in a cool, dry place out of direct sunlight. The ideal condition is 25C (77F). A range of 20C (68F) - 30C (86F) is acceptable. Test papers should not be stored in a refrigerator or freezer. Test papers should not be stored where they can be exposed to heat (for example, inside a vehicle or shed).

When stored properly, the Chlorine test papers have a minimum shelf-life of 3 years from the date of manufacture. Under proper storage conditions the strips will continue to perform properly beyond the expiration date.