



169 Lexington Court, Unit i
Waterloo, ON Canada N2J 4R9
Toll Free: (877) 746-4764
Local: (519) 746-4761
Email: info@indigoinstruments.com

CERTIFICATE of ANALYSIS

Product Code: [33813-Co](#)

Product Description: Cobalt Chloride Water Detection Test Strips.

These test strips are treated with a solution containing CoCl.

If moisture is present, the test strip will turn from blue to a light pink/white color. Exposure to 55 percent relative humidity or greater is sufficient enough to also cause a slow color change. If strips are heated or the surrounding air dries, the strips *can* revert to blue. If strips are exposed to sufficient water that the indicator chemical is washed out, the strips will not be able to be used further.

Each batch of test strips is tested before release for sale. A positive color change of batch master sample is required before the strips are approved for release.