



169 Lexington Court, Unit i
Waterloo, ON Canada, N2J 4R9
Toll Free: (877) 746-4764
Local: (519) 746-4761
Email: info@indigoinstruments.com

CERTIFICATION OF ANALYSIS

3-5-10 % HYDROGEN PEROXIDE TEST STRIP

The paper used to produce these Test Strips are impregnated with a 0.08% solution of chromogene indicator and inert ingredients (buffering salts, polymers and surfactant). There are no established health hazards, and the test strips are found to be stable (protect from sunlight and excessive heat until use).

The papers produced are tested against solutions of hydrogen peroxide 10%, 5% and 3% level of concentration with standard colorimetric assay verification and are verified to change color with corresponding lag-phase with a 99% frequency.

Prior to processing, the papers are air-dried and manufactured into strips then packaged into vials. This product is manufactured in the United States.

Expiration dates are 3 years from the date of packaging. The Test Strips should be kept at room temperature, in closed containers, out of direct sunlight.

The test strips are guaranteed to perform to specification for a minimum of 3 years from the date of manufacture (lot) date provided the test strips are stored in the original container and kept in a cool, dry environment out of direct sunlight.