



169 Lexington Court, Unit i
Waterloo, ON Canada, N2J 4R9
Toll Free: (877) 746-4764
Local: (519) 746-4761
Email: info@indigoinstruments.com

CERTIFICATION OF ANALYSIS

Product Code: [33818-PA2](#)

Product Description: Peracetic Acid Test Strips, 0-160 ppm

The PAA test strips consist of a chemically treated cellulose test pad (5mm²) adhered to a vinyl plastic backing. The paper is treated with a chemical mixture consisting of an indicator dye suitable for the range required, and various other chemicals – buffers, surfactants, etc.

Prior to release for sale, each lot of test strips is tested both for precision and accuracy. A verified color chart is used to evaluate the lot.

Precision Testing

- 10 strips from each lot are tested simultaneously in a PAA standard which has been verified using a secondary test kit method.
- Strips are evaluated on a pass/fail basis.
- All 10 strips must pass (little or no variation in color) for the lot to be acceptable.

Accuracy Testing

- PAA standards are prepared by serial dilution from a commercially available PAA concentrate.
- Standards are prepared at values of 0, 10, 30, 60, 80, and 160 ppm.
- Each standard is verified using a secondary test kit method.
- Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.
- Test strips are accepted if the color developed for the standard being evaluated visually falls within +/- ½ color unit.

The test strips are guaranteed to perform to specification for a minimum of 3 years from the date of manufacture (lot) date provided the test strips are stored in the original container and kept in a cool, dry environment out of direct sunlight.