




Sialkot Material Testing Laboratory

UNIT ESTABLISHED BY TRADE DEVELOPMENT AUTHORITY OF PAKISTAN MINISTRY OF COMMERCE GOVT. OF PAKISTAN

NEAR SUBLIME CHOWK, MARALA ROAD, SIALKOT-51310-PAKISTAN

TEL: 0092-52-3562818

FAX: 0092-52-3554217



Receipt No. 3060

PM/SIMTEL/2015/1967

March 4, 2015

ANALYTICAL/ TEST REPORT

In respect of : *M/s. Hyssop Surgico Sialkot*
Sample received from : *Representative*
Name of the Sample : *Mosquito Forceps 5"*
Grade / Type of the Material : *AISI-410*
Specification against which tested : **Analytical Results Got on Direct Emission Spectrometer**

Results are in percentage:

	Cr	12.498		
C	0.143		Mn	0.597
P	0.021		S	0.015
Si	0.271		Ni	0.102

This analytical/test report is based solely on the particular sample(s) supplied by the party and should not be reproduced in part.

Sampling has not been performed by the SIMTEL and the SIMTEL does not accept responsibility that the sample(s) supplied is/are truly representative sample(s) of any batch or stock or entire production.

While the SIMTEL agrees to take every reasonable precaution to ensure validity of its test results, it assumes no liability for the accuracy thereof beyond the amount of the fee charged for the analysis/test.

The party shall assume full responsibility for the ethical use of the results in the analysis/test reports and the laboratory shall be held free from any and all claims which may result from the use of such data by the party or others.

After completion of the report the sample will be preserved for three month until negotiated otherwise.

Sialkot
Material
Testing
Laboratory



Engr. Amjad Hussain
Project Director