



Sialkot Material Testing Laboratory

UNIT ESTABLISHED BY TRADE DEVELOPMENT AUTHORITY OF PAKISTAN MINISTRY OF COMMERCE GOVT. OF PAKISTAN

NEAR SUBLIME CHOWK, MARALA ROAD, SIALKOT-51310-PAKISTAN

TEL: 0092-52-3562818

FAX: 0092-52-3554217



PM/SIMTEL/2021/1968

March 4, 2021

Receipt No. 3060

ANALYTICAL/ TEST REPORT

In respect of : *M/s. Hyssop Surgico Sialkot*
Sample received from : *Representative*
Name of the Sample : *Dental Pick*
Grade / Type of the Material : *AISI-410*
Specification against which tested : ***Analytical Results Obtained on Direct Emission Spectrometer***

Results are in percentage:

Cr 12.534	
C 0.139	Mn 0.658
P 0.023	S 0.016
Si 0.273	Ni 0.101

This analytical/test report is based solely on the particular sample(s) supplied by the party and should not be reproduced in part.

Sampling has not been performed by the SIMTEL and the SIMTEL does not accept responsibility that the sample(s) supplied is/are truly representative sample(s) of any batch or stock or entire production.

While the SIMTEL agrees to take every reasonable precaution to ensure validity of its test results, it assumes no liability for the accuracy thereof beyond the amount of the fee charged for the analysis/test.

The party shall assume full responsibility for the ethical use of the results in the analysis/test reports and the laboratory shall be held free from any and all claims which may result from the use of such data by the party or others.

After completion of the report the sample will be preserved for three month until negotiated otherwise.



Engr. Amjad Hussain
Project Director